


NANOCOLOR® Phosphate 15




Method(e):

0801
0.30 - 15.00 mg/l P


0802
1.0 - 45.0 mg/l PO_4^{3-}

0803
0.7 - 34.5 mg/l P_2O_5


REF 985 080




Sample




1 x **NANOFIX R2**




Shake




120 °C / 30 min
100 °C / 1 h




Cool




1 x **NANOFIX R3**




200 µl **R4**



Shake




Clean



Measure

NANOCOLOR® Phosphate 15



Method(e):

0805

0.30 - 15.00 mg/l $\text{PO}_4\text{-P}$


0806

1.0 - 45.0 mg/l PO_4^{3-}


0804

0.7 - 34.5 mg/l P_2O_5


REF 985 080




Sample




1 x **NANOFIX R3**




200 µl **R4**



Shake



Clean



10'00 min



Measure